



THE AMERICAN ASSOCIATION FOR  
LABORATORY ACCREDITATION

## ACCREDITED LABORATORY

A2LA has accredited

**LABORATORY TESTING, INC.**

**Hatfield, PA**

for technical competence in the field of

### **Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).



Presented this 16<sup>th</sup> day of June 2009.

A handwritten signature in cursive script, reading "Peter Mlynar".

President  
For the Accreditation Council  
Certificate Number: 117.04  
Valid to March 31, 2011

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005  
& ANSI/NCSL Z540-1-1994

LABORATORY TESTING, INC.  
 2331 Topaz Drive  
 Hatfield, PA 19440  
 Frank Peszka Phone: 1 800 219 9095

CALIBRATION

Valid To: March 31, 2011

Certificate Number: 117.04

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	Best Uncertainty <sup>2,4</sup> ( $\pm$ )	Comments
Gage Blocks	Up to 1 in (>1 to 4) in (5 to 20) in	3.5 $\mu$ in (2.5 + 1L) $\mu$ in (7.7 + 0.4L) $\mu$ in	Mechanical comparison with reference blocks
Length Standards	(1 to 20) in (>20 to 40) in	(33 + 2.2L) $\mu$ in (48 + 2.3L) $\mu$ in	UMM
Calipers <sup>3</sup>	(0 to 24) in	(13 + 7L + 0.6R) $\mu$ in	Gage blocks
Outside Micrometers <sup>3</sup> –	(0 to 1) in (>1 to 6) in (>6 to 12) in (>12 to 24) in	60 $\mu$ in (51 + 9.4L) $\mu$ in (95 + 6L) $\mu$ in (12 + 13L) $\mu$ in	Gage blocks
Pitch Micrometers Disc Micrometers Interchangeable Anvil	(0 to 4) in (0 to 1) in (0 to 6) in	590 $\mu$ in 150 $\mu$ in 710 $\mu$ in	Optical comparator
Inside Micrometers	(0 to 40) in	160 $\mu$ in	UMM

Parameter/Equipment	Range	Best Uncertainty <sup>2,4</sup> ( $\pm$ )	Comments
Height Gages	(1 to 48) in	$(450 + 2.7L + 0.6R) \mu\text{in}$	Gage blocks surface plate
Indicators – Dial/Digital	Up to 2 in	$(28 + 0.6R) \mu\text{in}$	Indicator calibrator
Optical Comparators <sup>3</sup> – Linearity Magnification Angularity	(0 to 12) in (10X, 20X, 31.25X, 50X, 62.5X) 0° to 360°	$(650 + 0.6R) \mu\text{in}$ $(750 + 0.6R) \mu\text{in}$ 0°, 1', 17''	Accurite reticle and inspection scale
Pin Gages	Up to 1 in	23 $\mu\text{in}$	UMM, Pratt & Whitney Supermicrometer®
Plugs	Up to 3.0 in	11 $\mu\text{in}$	Federal comparator
Radius Gages	Up to 1 in	150 $\mu\text{in}$	Video measuring machine
Threaded Plugs – Major Diameter Pitch Diameter	Up to 3.5 in (3.5 to 6.5) in Up to 1.5 in (>1.5 to 3.5) in (>3.5 to 6.5) in	37 $\mu\text{in}$ 55 $\mu\text{in}$ 92 $\mu\text{in}$ 130 $\mu\text{in}$ 140 $\mu\text{in}$	UMM and Pratt & Whitney Supermicrometer®
Thread Wires (1-Wire)	Up to 0.26 in	24 $\mu\text{in}$	UMM and Pratt & Whitney Supermicrometer®
Plain Rings	Up to 3.0 in	13 $\mu\text{in}$	Comparator

Parameter/Equipment	Range	Best Uncertainty <sup>2,4</sup> ( $\pm$ )	Comments
Adjustable Thread Rings			
Minor Diameter	Up to 0.425 in (>0.425 to 2.8) in (>2.8 to 6.6) in	81 $\mu$ in 130 $\mu$ in 140 $\mu$ in	Bore gage Video measuring machine
Pitch Diameter	Up to 1.5 in (>1.5 to 3.5) in (>3.5 to 6.5) in	150 $\mu$ in 180 $\mu$ in 190 $\mu$ in	Set plug
Surface Plates <sup>3</sup>	16 to 195 in diagonal	(14 + 0.2L) $\mu$ in	Repeat-O-Meter Electronic levels
Pratt & Whitney Supermicrometer® / Bench Micrometers <sup>3</sup>	(0 to 2) in	20 $\mu$ in	Gage blocks
CMM <sup>3</sup> –			
Repeatability Linearity Bi-Directional Linearity Volumetric	Up to 1 in Up to 36 in 1” (6 to 18) in cube (18 to 40) in cube	51 $\mu$ in (59 + 10L) $\mu$ in 25 $\mu$ in (0 + 23L) $\mu$ in (1.4 + 23L) $\mu$ in	Sphere Step gage Gage blocks Ball bar

## II. Mechanical

Parameter/Equipment	Range	Best Uncertainty <sup>2,4</sup> ( $\pm$ )	Comments
Balances	Up to 1 lb (>1 to 5) lb (>5 to 10) lb (>10 to 25) lb (>1 to 50) lb (>50 to 100) lb (>100 to 200) lb (>200 to 500) lb (>500 to 1000) lb	(0.0012 + 0.6R) mg (20 + 0.6R) mg (25 + 0.6R) mg (110 + 0.6R) mg (230 + 0.6R) mg (510 + 0.6R) mg (720 + 0.6R) mg (1.2 + 0.6R) g (1.6 + 0.6R) g	Class F and class 1 weights  Class F weights

Parameter/Equipment	Range	Best Uncertainty <sup>2,4</sup> ( $\pm$ )	Comments
Balances (cont.)	(>1000 to 2000) lb  Up to 30 g (30 to 60) g (>60 to 120) g (>120 to 500) g	(2.25 + 0.6R) g  (0.02 + 0.6R) mg (0.03 + 0.6R) mg (0.05 + 0.6R) mg (0.07 + 0.6R) mg	Class F weights  Class 1 weights
Deadweight Testers	Up to 40 K psi	0.2 % of reading	Weights
Force	Up to 500 lbf (>500 to 100 000) lbf	1.2 g + 0.6R 0.35 % of reading	Dead weight, load cells
Mass	Up to 5 mg (>5 to 20) mg (>20 to 50) mg (>50 to 100) mg (>100 to 200) mg >200 mg to 5 g (>5 to 500) g >500 g to 5 kg	0.02 mg 0.0007 mg 0.21 mg 0.0056 mg 0.0023 mg 0.03 mg 0.08 mg 13 mg	Class 1 weights
Torque Wrenches	(0 to 50) in·lb (>50 to 250) in·lb (0 to 200) ft·lb (200 to 400) ft·lb (>400 to 750) ft·lb	0.86% fsd + 0.6R 0.54 % fsd + 0.6R 0.82% fsd + 0.6R 2.6 % fsd + 0.6R 0.7 % fsd + 0.6R	Torque calibrator
Indirect Verification of Rockwell Hardness Testers <sup>3</sup>	HR15N HR15T HR30N HRN30T HRN45N HRN45T	0.24 HR15N 0.31 HR15T 0.29 HR30N 0.33 HR30T 0.23 HR45N 0.40 HRN45T	Hardness blocks

Parameter/Equipment	Range	Best Uncertainty <sup>2</sup> (±)	Comments
Indirect Verification of Rockwell Hardness Testers <sup>3</sup> (cont.)	HRB High Mid Low  HRC High Mid Low	0.46 HRB 0.65 HRB 1.0 HRB  0.32 HRC 0.34 HRC 0.44 HRC	Hardness blocks
Pressure	(0 to 30) inHg (1 to 50) psi (50 to 500) psi (500 to 1000) psi (1 to 5) kpsi (5 to 40) kpsi	(0.07 + 0.6R) inHg (0.02 + 0.6R) psi (0.12 + 0.6R) psi (0.56 + 0.6R) psi (1.6 + 0.6R) psi 0.2 % of reading	Princo Princo Princo Heise Heise Cosa

<sup>1</sup> This laboratory offers commercial and field calibration service.

<sup>2</sup> “Best Uncertainty” is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. Best uncertainties represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The best uncertainty of a specific calibration performed by the laboratory may be greater than the best uncertainty due to the behavior of the customer’s device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the uncertainties achievable on a customer’s site can normally be expected to be larger than the Best Measurement Capabilities (BMC) that the accredited laboratory has been assigned as Best Uncertainty on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the calibration uncertainty being larger than the BMC.

<sup>4</sup> In the statement of best uncertainty,  $L$  is the numerical value of the nominal length of the device measured in inches;  $R$  is the numerical value of the resolution of the device in the indicated units.